




# Application of a long Memory FFT Analyzer in Speech Analysis

by N. Thrane,  
Bruel & Kjaer

*Paper presented at the Fourth F.A.S.E. Symposium on Acoustics and Speech, April 21-24, Venezia, Italy*



## Abstract

The introduction of dedicated FFT analyzers having a long input memory — i.e. 10K samples — has opened up new possibilities for simpler forms of speech analysis. Scanning a 1K sample long time window

through the 10K sample input memory permits seeing a running spectrum and studying its time variations. This can be repeated in its entirety or in part at will and at different speeds. Addition of a small desk-

top calculator allows shorter time windows to be used. Also a "scan" cepstrum can be performed for pitch extraction and formant analysis.

## Introduction

The analysis of speech as a non-stationary signal is normally implemented on mini-computers coupled to array processors. In contrast, this paper describes the use of a dedicated, self-contained instrument for the simpler forms of analysis. The instrument contains an ordinary FFT

processor which operates on a 1K sample time signal, thereby producing a 400 line spectrum ranging from DC up to a chosen maximum frequency. However, the input memory of the instrument is ten times longer than the length of the analysis window, i.e. 10K samples. This

allows the analysis window to be "scanned" through the total input memory and hence displaying the spectrum as it changes as a function of time. A more detailed discussion of the instrument can be found in the reference.

## Scan Analysis

When the instrument is used in the 0 — 5 kHz range the time length of the input memory is 800ms, which is sufficient to contain one or more syllables of speech. In this case the analysis window is 80ms long, but since a Hanning weighting function is normally used the effective length of the analysis window is only 30ms. A scan of this analysis window through the complete input memory can be performed either manually or automatically. In the automatic scan the step size between the indi-

vidual analyses can be varied from 80ms (i.e. 10 spectra per scan) down to 0,7ms (1153 spectra per scan). Each analysis takes approx. 100ms so the duration of a complete scan can be varied from 1s up to 2 minutes. In this way a "slow motion" presentation of the spectral changes can be observed directly on the instrument screen. Since the scan analysis does not change the original content of the input memory, the scan can be repeated at will, analysing any part of the signal at any convenient speed.

Fig.1 shows the result of such a scan analysis. The word "Montreal" was recorded in the 800ms input memory; however, only the part containing the "ea" sound was analysed. The step size used was 3,1ms. The three-dimensional plot used here to present the results does not reflect the visual impression of the live "slow motion" analysis, but only the complexity of the results.




Fig.1. Scan spectrum of "ea" in "Montreal"

## Cepstrum Liftering

In order to simplify the information contained in the 3-D landscape of Fig.1 a "cepstrum liftering" was performed of each spectrum. An example is shown in Fig.2. The log

power spectrum of a vowel sound is seen in Fig.2.a and its cepstrum in Fig.2.b. (The cepstrum is defined as the inverse Fourier transform of the log power spectrum).

The rapid fluctuations in the spectrum caused by the voice frequency (150Hz) and its harmonics are reflected in the cepstrum as a sharp peak at a "quefrequency" of 6,6ms

plus the second and third "harmonics". Conversely, the slowly varying parts of the spectrum, i.e. the formants, are found at low quefrequencies. Since the two effects are additive and separated in quefrequency, application of the "short pass lifter" characteristics shown in Fig.2.c will remove the influence of the harmonics of the voice from the cepstrum. A smooth lifter characteristic, being half of a Hanning function of length corresponding to the voice quefrequency, was used to avoid introduction of discontinuities in the liftered cepstrum. Finally, the short pass liftered cepstrum was Fourier transformed back to the frequency domain, producing the log power spectrum shown in Fig.2.d, where the formant characteristics are clearly seen.

The cepstrum liftering procedure was performed using a small desktop calculator in conjunction with the FFT analyser, both communicating over their common IEC data bus.

The final result is shown in Fig.3 as a 3-dimensional plot. In contrast




Fig.2. Cepstrum liftering. a) log power spectrum of vowel. b) magnitude of cepstrum. c) short pass lifter characteristic. d) short pass liftered log power spectrum

to the analysis shown in Fig.1, the cepstrum liftering has emphasised the formant structures so that the

development with time can now easily be followed and studied in detail.




Fig.3. Short pass liftered scan spectrum of "ee" in "Montreal"

## Conclusion

An FFT analyser having a long input memory can be used for a live "slow motion" scan analysis of speech. In conjunction with a small desktop calculator it provides a number of more refined analysis meth-

ods: scan analysis of short pass filtered spectra, scan of voice pitch either in the frequency or in the quency domain, and scan analysis of signals changing very rapidly by use of time windows shorter

than those normally available.

### Reference

N. Thrane: Zoom, FFT, Technical Review No. 2, 1980, Brüel & Kjær.

The instrument described in this paper is the Brüel & Kjær Type 2033, High Resolution Signal Analyzer.

A.V. Oppenheimer & R.W. Schafer, Digital Signal Processing, Prentice Hall, N.J., 1975.


R.B. Randall, and J. Hee, "Cepstrum Analysis", Technical Review, Brüel & Kjær, to be published.

Further information on the cepstrum technique and its applications can be found in the following references:

D.G. Childers, D.P. Skinner, and R.C. Kemerait, "The Cepstrum: A guide to processing", Proc. IEEE, vol. 65, pp. 1428—1443.

# The Unique Approach to Speech Analysis ....BRÜEL & KJÆR SCAN

Brüel & Kjær's unique High Resolution Signal Analyzer 2033 operates on 10240 samples of the time function. That means it can preserve 800 msec. of signal on a 5000 Hz range, which could represent three or four syllables of speech.



The Unique Scan function is then available to give a "live" display of a slow motion analysis of the changing spectral content of the speech signal, with respect to time, with selectable scan rates to suit the user.

The Brüel & Kjær High Resolution Signal Analyzer Type 2033.

80-188



**Brüel & Kjær Instruments, Inc.**